Table 1-1. Model HP 8340B Specifications and Supplemental Performance Characteristics (1 of 15)

## NOTE

Specifications are the performance standards, or limits, against which the instrument may be tested. The following Specifications apply for temperatures between 0 and +50°C except were noted. Specifications apply with the PEAK function ON in the CW and MANUAL modes of operation, and with periodic use of AUTO TRACKING CALIBRATION in swept operation.

Supplemental Performance Characteriaitcs are in *Italics* in this table and are intended to provide information useful in applying the instrument by giving typical, but non-warranted, performance parameters. These are denoted as "typical," "nominal," or "approximate."

# FREQUENCY

## CW MODE

Range: 0.01 to 26.5 GHz

**Resolution:**  $n \times 1$  Hz

Where N = harmonic multiplication (1 to 4). Refer to Frequency ranges and Bandswitch Points description below.

Accuracy: Same as Time Base Accuracy

## Time Base

Frequency: 10 MHz

## Aging Rate:

 $1 \times 10^{-9}$  per day,  $2.5 \times 10^{-7}$  per year after 72 hour warm up if HP 8340B has been disconnected from ac power for less than 24 hours. Aging rate is achieved after 7 to 30 days warm up if HP 8340B has been disconnected from ac power for greater than 24 hours.

**Temperature Coefficient:** Typically  $< 1 \times 10^{-10}$  per °C

Change due to 10% line voltage change: Typically  $< 1 \times 10^{-n}$ 

Accuracy:

**Overall accuracy** of internal time base is a function of time base calibration  $\pm$  aging rate  $\pm$  temperature effects  $\pm$  line effects.

Switching Time: <50 msec (PEAK function off) Fast Phase Lock Mode reduces typical switching time to <20 msec.)

# CENTER FREQUENCY/SWEEP WIDTH MODE (CF/ΔF)

Range: 10.00005 MHz to 26.49999995 GHz (center frequency) 100 Hz to 26.49 GHz (sweep width)

**Resolution:** Approximately 0.1% of sweep width ( $\Delta F$ )

Readout Accuracy with respect to sweep out voltage (sweep time > 100 msec):

 $\Delta \leq 0 \times 5$  MHz:  $\pm$  1% of indicated sweep width ( $\Delta F$ )  $\pm$  time base accuracy

 $\Delta >$  n  $\times$  5 MHz to < 300 MHz:  $\pm$  2% of indicated sweep width ( $\Delta$ F)

 $\Delta \ge$  300 MHz: ±1% of indicated sweep width ( $\Delta$ F), or ± 50 MHz, whichever is less.

Where n = harmonic multiplication number (1 to 4). Refer to Frequency Ranges and Bandswitch Points description below.

"Time Base effects Center Frequency accuracy only, not sweep width accuracy.

Table I-I. Model HP 8340B Specifications and Supplemental Performance Characteristics (2 of 15)

# FREQUENCY (Cont'd)

#### START/STOP MODE

## Range

Start: 10 MHz to 26 499.9999 MHz Stop: 10.0001 MHz to 26.5 GHz

Resolution: Typically, the same as Center Frequency/Sweep Width mode:

Readout Accuracy with respect to sweep out voltage (sweep time > 100 msec): Same as Center Frequency/Sweep Width Mode.

#### FREQUENCY MARKERS

All 5 markers are independently variable and have the same specifications.

Range: 10 MHz to 26.5 GHz

**Resolution:** Approximately 0.1% of sweep width  $(\Delta F)$ 

Readout Accuracy (sweep time >100 msec): Same as CENTER FREQUENCY/SWEEP WIDTH MODE (CF/ $\Delta$ F).

"Time base accuracy is not a factor in MKRA Mode.

#### FREQUENCY RANGES AND BANDSWITCH POINTS

For bands 0 and 1, the HP 8340B's output is derived from the fundamental frequency of its internal 2.3 to 7.0 GHz YIG-tuned oscillator (n = 1). For bands 2, 3, and 4 the output is derived from the 2nd, 3rd, or 4th harmonic of the oscillator (n = 2, 3, or 4).

Bandswitch points in CW Mode (only) always occur at the following points:

Band 0 to 1: 2.3 GHz Band 1 to 2: 7.0 GHz Band 2 to 3: 13.5 GHz Band 3 to 4: 20.0 GHz

Bandswitch points in each of the swept modes (CF/ΔF, START/STOP) and the MANUAL SWEEP mode normally occur at the following points (with the exception listed below):

Band 0 to 1: 2.4 GHz Band 1 to 2: 7.0 GHz Band 2 to 3: 13.5 GHz Band 3 to 4: 20.0 GHz

The swept mode bandswitch points are illustrated in Figure 1.





Note that the bands overlap. However, any sweep will be executed with the minimum number of bandswitch points. If the start frequency is above the lower limit for a given band, the sweep will start in that band and not the next lower one. If the stop frequency exceeds the upper limit of a given band by an amount greater than  $0.004 \times \Delta F$ , a bandswitch will occur at that band's upper limit.

# SPECTRAL PURITY

(Spectral Purity specifications apply for CW mode and all swept modes, unless otherwise stated.)

SPURIOUS SIGNALS	Bands and Approximate Frequency Ranges (GHz) (See Frequency Ranges and Bandswitch Points for complete description)					
(Expressed in d8 relative to the carrier level (dBc) at ALC level of 0dBm)	Band 0 0.01 to <2.3	<b>Band 1</b> 2.3 to <7.0	Band 2 7.0 to <13.5	Band 3 13.5 to <20.0	Band 4 20.0 to 26.5	
Harmonics (only up to 26.5 GHz)	<-35	<-35	<-35	<-35	<-35	
Subharmonics and multiples theref (up to 26.5 GHz)		-	<-25	<-25	<-20	
Non-harmonically related supurious (CW and Manual Sweep mode only)	<-50	<-70	<-64	<-60	<-58	

Table 1-1. Mor	del HP 8340B Specifications and	Supplemental Performance	Characteristics (4 of 15)
----------------	---------------------------------	--------------------------	---------------------------

	SPECTR		(Cont'd)		
SPURIOUS SIGNALS (Cont'd)					
Power line related and spi	urious due to fan r	rotation within 5 l	Hz below line free	quency, and multi	ples thereof
(CW mode only, all power	Band 0 Band 1		Band 2	Band 3	Band 4
levels)	0.01 to <2.3	2.3 to <7.0	7.0 to <13.5	13.5 to <20.0	-20.0 to 26.
Offset <300 Hz from carrier	<-50	<-50	<-44	<-40	<-38
Offset 300 Hz to 1 kHz from				- 50	
carrier	<-60	<-60	<-54	<-50	<-48
Offset >1 kHz from carrier	<-65	<-65	<-59	< -55	< - 53
SINGLE-SIDEBAND PHASE NO	)ISE (dBc/1 Hz BW	, CW Mode, all p	ower levels)		
STANDARD INSTRUMENT			-		
Offset 30 Hz from carrier	<−64	<-64	<-58	< - 54	<-52
Offset 100 Hz from carrier	<-70	<-70	<-64	<-60	<-58
Offset 1 kHz from carrier	<-78	<78	<-72	< -68	<-66
Offset 10 kHz from carrier	<-86	<-86	<-80	<76	<-74
Offset 100 kHz from carrier	<-107	<-107	<-101	<-97	<-95
OPTION 007, Relaxed Phase Noise Specifications					
Offset 100 Hz from carrier	<-67	<-67	<61	<-57	< - 55
Offset 1 kHz from carrier	<-75	<-75	<-69	<-65	<-63
Offset 10 kHz from carrier	<-83	<-83	<-77	<-73	<-71
Offset 100 kHz from carrier	<-107	<-107	<~101	<-97	<-95

TYPICAL FREQUENCY STABILITY, 50 Hz - 15 kHz post detection bandwidth

Typical Residual FM in CW Mode: <  $n \times 60~Hz~rms$ 

Typical Residual FM in Swept Mode:  $\Delta F > n \times 5 \text{ MHz}$ :  $< n \times 25 \text{ kHz rms}$  $\Delta F \leq n \times 5 \text{ MHz}$ : Same as CW mode

 $\Delta r \leq n \times 5$  MHZ. Same as CW mode

Where n = harmonic multiplication number (1 to 4). Refer to Frequency Ranges and Bandswitch Points description above.

Table 1-1. Model HP 8340B Specifications and Supplemental Performance Characteristics (5 of 15)

		RF OL	JTPUT				
MAXIMUM LEVELED	(See	Bands and Approximate Frequency Ranges (GHz) (See Frequency Ranges and Bandswitch Points for complete description)					
POWER	Band 0	Band 1	Band 1 Band 2		Band 4		
	0.01 to <2.3	2.3 to <7.0	7.0 to <13.5	13.5 to <20.0	20.0 to <23.0	23.0 to 26.5	
STANDARD INSTRUMENT	+10.0 dBm	+12.0 dBm	+10.0 dBm	9.0 dBm	+3.0 dBm	+1.0 dBm	
OPTION 001 (F.P. Out w/o Atten.)	+10.0 dBm	+ 13.0 dBm	~12.0 dBm	+11.0 dBm	+6.0 dBm	+4.0 dBm	
OPTION 004 (R.P. Out w/Atten.)	÷ 10.0 dBm	+11.0 dBm	+9.0 dBm	+7.0 dBm	-+1.0 dBm		
OPTION 005 (R.P. Out w/o Atten.)	+ 10.0 dBm	+12.0 dBm	+11.0 dBm	+9.0 dBm	+4.0 d8m		
OPTION 001 (F.P. Out RF OFF When the RF key is tur				20 dBm			
reduced to a level $< -$			рау мів геао О	FF and a 0 dBn	n signal will typ	ically be	
reduced to a level < - OUTPUT POWER RESOLUT "ENTRY DISPLAY": (			· · ·	"Display: 0.1		ically be	
OUTPUT POWER RESOLU" "ENTRY DISPLAY": (	100 dBm. FION 0.05 dB	·······	"POWER dBm s and Approxin	" Display: 0.1	dB • Ranges (GHz)		
OUTPUT POWER RESOLU" "ENTRY DISPLAY": (	100 dBm. FION 0.05 dB	Band	"POWER dBm s and Approxin Ranges and Ba B	" Display: 0.1	dB Ranges (GHz) s for complete d Ba		
OUTPUT POWER RESOLU "ENTRY DISPLAY": ( OUTPUT POWER ACCURA STANDARD INSTRUMENT	100 dBm. FION 0.05 dB CY2	Band (See Frequency Band O	"POWER dBm s and Approxin Ranges and Ba B	" Display: 0.1 nate Frequency ndswitch Points ands 1 - 3 3 to <20.0	dB Ranges (GHz) s for complete d Ba 20.0	description) and 4 to 26.5	
OUTPUT POWER RESOLU" "ENTRY DISPLAY": ( OUTPUT POWER ACCURA STANDARD INSTRUMENT +18 to +10 dBm3	100 dBm. FION 0.05 dB CY2	Band (See Frequency Band 0 0.01 to <2.3	"POWER dBm s and Approxin Ranges and Ba B	" Display: 0.1 nate Frequency ndswitch Points ands 1 - 3 3 to <20.0 ±1.8 dB	dB Ranges (GHz) s for complete d <b>Ba</b> 20.0	description) and 4 to 26.5 2.3 dB	
OUTPUT POWER RESOLU" "ENTRY DISPLAY": ( DUTPUT POWER ACCURA STANDARD INSTRUMENT +18 to +10 dBm <sup>3</sup> +10 to -9.95 dBm	100 dBm. FION 0.05 dB CY2	Band (See Frequency Band 0 0.01 to < 2.3 	"POWER dBm s and Approxin Ranges and Ba B	" Display: 0.1 nate Frequency indswitch Points ands 1 - 3 3 to <20.0 ±1.6 dB ±1.5 dB	dB r Ranges (GHz) s for complete d <b>Bs</b> 20.0 = 2 ± 2	description) and 4 to 26.5 2.3 dB 2.0 dB	
OUTPUT POWER RESOLUT "ENTRY DISPLAY": ( DUTPUT POWER ACCURA STANDARD INSTRUMENT +18 to +10 dBm <sup>3</sup> +10 to -9.95 dBm -10 to -19.95 dBm	100 dBm. FION 0.05 dB CY2	Band (See Frequency Band 0 0.01 to < 2.3  ± 0.9 dB ± 1.2 dB	"POWER dBm s and Approxin Ranges and Ba B	" Display: 0.1 nate Frequency indswitch Points ends 1 - 3 3 to <20.0 ± 1.6 dB ± 1.5 dB ± 2.0 dB	dB Ranges (GHz) s for complete d 20.0 = 2 ± 2 ± 2	) description) and 4 to 26.5 2.3 dB 2.0 dB 2.5 dB	
OUTPUT POWER RESOLUT "ENTRY DISPLAY": ( OUTPUT POWER ACCURA STANDARD INSTRUMENT +18 to +10 dBm <sup>3</sup> +10 to -9.95 dBm -10 to -19.95 dBm -20 to -49.95 dBm	100 dBm. FION 0.05 dB CY2	Band (See Frequency Band 0 0.01 to < 2.3 ± 0.9 dB ± 1.2 dB ± 1.5 dB	"POWER dBm s and Approxin Ranges and Ba B	" Display: 0.1 nate Frequency indswitch Points ands 1 - 3 3 to <20.0 ±1.8 dB ±1.5 dB ±2.0 dB ± 2.3 dB	dB r Ranges (GHz) s for complete ( 20.0 = 2 ± 2 ± 2 ± 2 ± 2 ± 2	) description) and 4 to 26.5 2.3 dB 2.3 dB 2.5 dB 2.8 dB	
OUTPUT POWER RESOLU "ENTRY DISPLAY": ( OUTPUT POWER ACCURA STANDARD INSTRUMENT +18 ta +10 dBm <sup>3</sup> +10 to -9.95 dBm -10 to -19.95 dBm -20 to -49.95 dBm -50 to -79.95 dBm	100 dBm. FION 0.05 dB CY2	Band (See Frequency Band 0 0.01 to < 2.3 ± 0.9 dB ± 1.2 dB ± 1.2 dB ± 1.8 dB	"POWER dBm s and Approxin Ranges and Ba B	" Display: 0.1 mate Frequency indswitch Points ands 1 - 3 3 to <20.0 ± 1.8 dB ± 1.5 dB ± 2.0 dB ± 2.3 dB ± 2.6 dB	dB Ranges (GHz) s for complete of 20.0 = 2 ± 2 ± 2 ± 2 ± 3	description) and 4 to 26.5 2.3 dB 2.0 dB 2.5 dB 2.8 dB 3.1 dB	
OUTPUT POWER RESOLUT "ENTRY DISPLAY": ( OUTPUT POWER ACCURA STANDARD INSTRUMENT +18 to +10 dBm <sup>3</sup> +10 to -9.95 dBm -10 to -19.95 dBm -20 to -49.95 dBm	100 dBm. FION 0.05 dB	Band (See Frequency Band 0 0.01 to < 2.3 ± 0.9 dB ± 1.2 dB ± 1.5 dB	"POWER dBm s and Approxin Ranges and Ba B	" Display: 0.1 nate Frequency indswitch Points ands 1 - 3 3 to <20.0 ±1.8 dB ±1.5 dB ±2.0 dB ± 2.3 dB	dB Ranges (GHz) s for complete of 20.0 ±2 ±2 ±2 ±2 ±3 ±3	) description) and 4 to 26.5 2.3 dB 2.3 dB 2.5 dB 2.8 dB	
OUTPUT POWER RESOLUT "ENTRY DISPLAY": (           OUTPUT POWER ACCURA           STANDARD INSTRUMENT +18 ta +10 dBm <sup>3</sup> +10 to -9.95 dBm -10 to -19.95 dBm -20 to -49.95 dBm -20 to -49.95 dBm -50 to -79.95 dBm -80 to -100 dBm -100 to -110 dBm (typic OPTION 004 (Rear Panel Output w/Atter	100 dBm. FION 0.05 dB CY2	Band (See Frequency Band 0 0.01 to < 2.3 	"POWER dBm s and Approxin Ranges and Ba B	" Display: 0.1 mate Frequency indswitch Points ands 1 - 3 3 to <20.0 ± 1.6 dB ± 1.5 dB ± 2.0 dB ± 2.3 dB ± 2.6 dB ± 2.9 dB ± 3.7 dB	dB Ranges (GHz) a for complete a 20.0	description) and 4 to 26.5 2.3 dB 2.0 dB 2.5 dB 2.8 dB 3.1 dB 3.4 dB 4.2 dB	
OUTPUT POWER RESOLUT "ENTRY DISPLAY": (           OUTPUT POWER ACCURA           STANDARD INSTRUMENT +18 ta +10 dBm <sup>3</sup> +10 to -9.95 dBm -10 to -19.95 dBm -20 to -49.95 dBm -20 to -49.95 dBm -50 to -79.95 dBm -80 to -100 dBm -100 to -110 dBm (typic OPTION 004 (Rear Panel Output w/Atte +18 to +10 dBm <sup>3</sup>	100 dBm. FION 0.05 dB CY2	Band (See Frequency Band 0 0.01 to <2.3 ±0.9 dB ±1.2 dB ±1.2 dB ±1.5 dB ±1.5 dB ±2.1 dB ±2.9 dB	"POWER dBm s and Approxin Ranges and Ba B	" Display: 0.1 mate Frequency indswitch Points ands 1 - 3 3 to <20.0 ± 1.6 dB ± 1.5 dB ± 2.0 dB ± 2.3 dB ± 2.6 dB ± 2.9 dB ± 3.7 dB ± 2.0 dB	dB r Ranges (GHz) s for complete of 20.0	description) and 4 to 26.5 2.3 dB 2.0 dB 2.5 dB 2.8 dB 3.1 dB 3.4 dB 4.2 dB 4.2 dB	
OUTPUT POWER RESOLUT "ENTRY DISPLAY":           OUTPUT POWER ACCURA           STANDARD INSTRUMENT +18 ta +10 dBm <sup>3</sup> +10 to -9.95 dBm -10 to -19.95 dBm -20 to -49.95 dBm -20 to -49.95 dBm -50 to -79.95 dBm -80 to -100 dBm -100 to -110 dBm (typic OPTION 004 (Rear Panel Output w/Atte +18 to +10 dBm <sup>3</sup> +10 to -11.95 dBm	100 dBm. FION 0.05 dB CY2	Band (See Frequency Band 0 0.01 to <2.3 ±0.9 dB ±1.2 dB ±1.2 dB ±1.5 dB ±1.5 dB ±2.1 dB ±2.9 dB ±1.0 dB	"POWER dBm s and Approxin Ranges and Ba B	" Display: 0.1 mate Frequency indswitch Points ands 1 - 3 3 to <20.0 ± 1.6 dB ± 1.5 dB ± 2.0 dB ± 2.6 dB ± 2.9 dB ± 3.7 dB ± 2.0 dB ± 1.7 dB	dB r Ranges (GHz) s for complete ( 20.0	description) and 4 to 26.5 2.3 dB 2.0 dB 2.5 dB 2.8 dB 3.1 dB 3.4 dB 4.2 dB 2.5 dB 2.5 dB 2.2 dB	
OUTPUT POWER RESOLUT "ENTRY DISPLAY":           OUTPUT POWER ACCURA           STANDARD INSTRUMENT +18 ta +10 dBm <sup>3</sup> +10 to -9.95 dBm -10 to -19.95 dBm -20 to -49.95 dBm -50 to -79.95 dBm -60 to -100 dBm -100 to -110 dBm (typic OPTION 004 (Rear Panel Output w/Atte +18 to -10 dBm <sup>3</sup> +10 to -11.95 dBm -12 to -21.95 dBm	100 dBm. FION 0.05 dB CY2	Band (See Frequency Band 0 0.01 to <2.3 ±0.9 dB ±1.2 dB ±1.2 dB ±1.5 dB ±1.5 dB ±2.1 dB ±2.9 dB ±2.9 dB	"POWER dBm s and Approxin Ranges and Ba B	" Display: 0.1 nate Frequency indswitch Points ands 1 - 3 3 to <20.0 ± 1.8 dB ± 1.5 dB ± 2.0 dB ± 2.3 dB ± 2.6 dB ± 2.9 dB ± 3.7 dB ± 1.7 dB ± 2.2 dB	dB r Ranges (GHz) a for complete ( 20.0	description) and 4 to 26.5 2.3 dB 2.0 dB 2.5 dB 2.8 dB 3.4 dB 4.2 dB 2.5 dB 2.2 dB 2.2 dB 2.7 dB	
OUTPUT POWER RESOLUT "ENTRY DISPLAY":           OUTPUT POWER ACCURA           STANDARD INSTRUMENT +18 ta +10 dBm <sup>3</sup> +10 to -9.95 dBm -10 to -19.95 dBm -20 to -49.95 dBm -50 to -79.95 dBm -50 to -79.95 dBm -80 to -100 dBm -100 to -110 dBm (typic OPTION 004 (Rear Panel Output w/Atte +18 to +10 dBm <sup>3</sup> +10 to -11.95 dBm -12 to -21.95 dBm -22 to -51.95 dBm	100 dBm. FION 0.05 dB CY2	Band (See Frequency Band 0 0.01 to <2.3 ±0.9 dB ±1.2 dB ±1.2 dB ±1.5 dB ±1.5 dB ±2.1 dB ±2.9 dB ±1.3 dB ±1.3 dB ±1.3 dB ±1.6 dB	"POWER dBm s and Approxin Ranges and Ba B	" Display: 0.1 nate Frequency indswitch Points ands 1 - 3 3 to <20.0 ± 1.8 dB ± 1.5 dB ± 2.0 dB ± 2.9 dB ± 2.9 dB ± 3.7 dB ± 2.2 dB ± 2.2 dB ± 2.5 dB	dB r Ranges (GHz) a for complete ( 20.0	description) and 4 to 26.5 2.3 dB 2.0 dB 2.5 dB 2.8 dB 3.1 dB 3.4 dB 4.2 dB 4.2 dB 2.2 dB 2.2 dB 2.7 dB 3.0 dB	
OUTPUT POWER RESOLUT "ENTRY DISPLAY": ( OUTPUT POWER ACCURA STANDARD INSTRUMENT +18 ta +10 dBm <sup>3</sup> +10 to -9.95 dBm -10 to -19.95 dBm -20 to -49.95 dBm -50 to -79.95 dBm -50 to -79.95 dBm -80 to -100 dBm -100 to -110 dBm (typic OPTION 004 (Rear Panel Output w/Atte +18 to +10 dBm <sup>3</sup> +10 to -11.95 dBm -12 to -21.95 dBm -22 to -51.95 dBm -52 to -81.95 dBm	100 dBm. FION 0.05 dB CY2	Band (See Frequency Band 0 0.01 to < 2.3 ± 0.9 dB ± 1.2 dB ± 1.2 dB ± 1.5 dB ± 2.1 dB ± 2.9 dB ± 2.9 dB ± 1.3 dB ± 1.3 dB ± 1.6 dB ± 1.9 dB	"POWER dBm s and Approxin Ranges and Ba B	" Display: 0.1 nate Frequency indswitch Points ands 1 - 3 3 to <20.0 ± 1.8 dB ± 1.5 dB ± 2.0 dB ± 2.3 dB ± 2.9 dB ± 2.9 dB ± 3.7 dB ± 2.2 dB ± 2.2 dB ± 2.5 dB ± 2.8 dB	dB r Ranges (GHz) a for complete (	) description) and 4 to 26.5 2.3 dB 2.0 dB 2.5 dB 2.8 dB 3.1 dB 3.4 dB 4.2 dB 2.5 dB 2.2 dB 2.7 dB 3.0 dB 3.3 dB	
OUTPUT POWER RESOLUT "ENTRY DISPLAY": ( OUTPUT POWER ACCURA STANDARD INSTRUMENT +18 ta +10 dBm <sup>3</sup> +10 to -9.95 dBm -10 to -19.95 dBm -20 to -49.95 dBm -50 to -79.95 dBm -80 to -100 dBm -100 to -110 dBm (typic OPTION 004 (Rear Panel Output w/Atte +18 to +10 dBm <sup>3</sup> +10 to -11.95 dBm -12 to -21.95 dBm -22 to -51.95 dBm	100 dBm. FION 0.05 dB CY2	Band (See Frequency Band 0 0.01 to <2.3 ±0.9 dB ±1.2 dB ±1.2 dB ±1.5 dB ±1.5 dB ±2.1 dB ±2.9 dB ±1.3 dB ±1.3 dB ±1.3 dB ±1.6 dB	"POWER dBm s and Approxin Ranges and Ba B	" Display: 0.1 nate Frequency indswitch Points ands 1 - 3 3 to <20.0 ± 1.8 dB ± 1.5 dB ± 2.0 dB ± 2.9 dB ± 2.9 dB ± 3.7 dB ± 2.2 dB ± 2.2 dB ± 2.5 dB	dB Ranges (GHz) a for complete of 8 20.0	description) and 4 to 26.5 2.3 dB 2.0 dB 2.5 dB 2.8 dB 3.1 dB 3.4 dB 4.2 dB 4.2 dB 2.2 dB 2.2 dB 2.7 dB 3.0 dB	

RF OUTPUT (Cont'd)				
OUTPUT POWER ACCURACY (Cont'd) OPTION 001	Band 0 0.01 to <2.3	Bands 1-3 2.3 to <20.0	Band 4 20.0 to 26.5	
(Front Panel Output w/o Attenuator) +18 to +10 dBm <sup>3</sup> +10 to -10 dBm -10 to -20 dBm	±0.9 dB ±1.7 dB	±1.6 dB ±1.3 dB ±2.1 dB	±2.0 dB ±1.7 dB ±2.5 dB	
OPTION 005 (Rear Panel Output w/o Attenuator)			· -	
+18 to +10 dBm³ −10 to -10 dBm −10 to -20 dBm		= 1.8 dB ± 1.5 dB ± 2.3 dB	±2.2 dB ± 1.9 dB ±2.7 dB	
FLATNESS (internally leveled)			,	
STANDARD INSTRUMENT +18 to +10 dBm <sup>3</sup> +10 to -9.95 dBm -10 to -19.95 dBm -20 to -49.95 dBm -50 to -79.95 dBm -80 to -100 dBm -100 to -110 dBm (typical)	± 0.6 dB ± 0.9 dB ± 1.2 dB ± 1.4 dB ± 1.7 dB = 1.9 d8	± 1.2 dB ± 1.1 dB ± 1.6 dB ± 1.9 dB ± 2.2 dB ± 2.5 dB ± 3.1 d9	± 1.7 dB ± 1.6 dB ± 2.1 dB ± 2.4 dB ± 2.7 dB ± 3.0 dB = 3.6 dB	
OPTION 004 (Rear Panel Output w/Attenuator) +18 to -10 dBm <sup>3</sup> +10 to -11.95 dBm -12 to -21.95 dBm -22 to -51.95 dBm -52 to -81.95 dBm -82 to -100 dBm -100 to -110 dBm (typical)	± 0.7 dB ± 1.0 dB ± 1.3 dB ± 1.5 dB ± 1.8 dB ± 2.0 d8	± 1.4 dB ± 1.3 dB ± 1.8 dB ± 2.1 dB ± 2.4 dB ± 2.7 dB ± 3.3 d9	± 1.9 dB ± 1.8 dB ± 2.3 dB = 2.6 dB = 2.9 dB ± 3.2 dB ± 3.8 dB	
DPTION 001 (Front Panel Output w/o Attenuator) +18 to +10 dBm <sup>3</sup> +10 to -10 dBm -10 to -20 dBm	±0.6 dB ±0.8 dB	± 1.0 dB ± 0.9 dB ± 1.5 dB	± 1.4 dB ± 1.3 dB ± 1.9 dB	
OPTiON 005 (Rear Panel Output w/o Attentuator) +18 to +10 dBm <sup>3</sup> +10 to -10 dBm -10 to -20 dBm	±0.7 dB ±0.9 dB	±1.2 dB ±1.1 dB ±1.7 dB	±1.6 dB ±1.5 dB ±2.1 dB	

 Table 1-1. Model HP 8340B Specifications and Supplemental Performance Characteristics (6 of 15)





### **RF OUTPUT CONNECTOR**

Output Impedance: Nominally 50 Ohms

Typical Source SWR (Internally leveled only): 0.1 to <2.3 GHz: *Typically <1.3:1* 2.3 to <18.0 GHz: *Typically <1.6:1* 18.0 to 26.5 GHz: *Typically <2.0:1* 

STABILITY WITH TEMPERATURE: Typically ±0.01 dB/°C

# OUTPUT LEVEL SWITCHING TIME:

Typically <10 ms to be within 0.1 dB of final value with no attenuator range change (internally leveled only).

# POWER SWEEP

#### Range:

Displayed: 0 to 40 dB/sweep

Actual: At least 10 dB at any given frequency (at least 20 dB in DECOUPLED mode: see Figure 3 below).

Resolution: 0.05 d8/sweep

Accuracy:

Starting Power Level: Same as Output Power Accuracy Power Sweep Width and Linearity: See Figure 2



Table 1-1. Model HP 8341B Specifications and Supplemental Performance Characteristics (8 of 15)

 Table 1-1.
 Model HP 8340B Specifications and Supplemental Performance Characteristics (9 of 15)

# PULSE MODULATION

Not Applicable to HP 8340Bs Equipped with Option 006, Delete Pulse Modulation. (CW mode, and all specifications are typical for frequencies <400 MHz)

ON/OFF RATIO: >80 dB

RISE (T<sub>R</sub>) AND FALL (T<sub>F</sub>) TIMES:  $\leq$ 25 nanoseconds

MINIMUM INTERNALLY LEVELED RF PULSE WIDTH (T<sub>RF</sub>): 100 nanoseconds

MINIMUM UNLEVELED RF PULSE WIDTH: Typically 25 nanoseconds

## PULSE REPETITION FREQUENCY (PRF)

Non-leveled operation (SHIFT METER): Typically dc to 20 MHz.

Internally leveled operation: 100 Hz to 5 MHz (typically 100 Hz to 500 kHz for RF frequencies <400 MHz).

MAXIMUM PEAK POWER: Same as specified maximum leveled power. (See RF OUTPUT).

ACCURACY OF INTERNALLY LEVELED RF PULSE Vp (relative to CW mode level):

(Note that the ALC attempts to hold pulse amplitude to save level as leveled CW signal.)

Pulse Width	Bai	Band G			
	0.01 to 0.4	0.4 to <2.3	2.3 to 26.5		
100 to <200 ns	_	+3/0.3 dB*	+1.5/-0.3 dB		
200 to <500 ns	i —	+1.5/-0.3 dB*	<u>-</u> 0.3 dB		
≥500 ns	_	±0.3 dB	± 0.3 dB		
† to <2 μs	+3/~0.3 dB	· —	_		
2 to <5 μs	+1.5/-0.3 dB		_		
≥5 μs	±0.3 dB		_		

## SIMULTANEOUS AM AND PULSE (Peremeters shown are typical)

# AM BANOWIDTH AT 30% DEPTH

DC coupled, typical 3 dB point:

Unleveled (Shift Meter)*
10 <b>0 kHz</b>
J.5 µsec





TTL compatible. (Open circuit is TTL high level and keeps RF on.) Damage level is +12 Vdc, -20 Vdc. Refer to Section III, Operation, for input circuit diagram.



 Table 1-1. Model HP 8340B Specifications and Supplemental Performance Characteristics (11 of 15)



MODULATION RATE: 50 kHz to 10 MHz (3 dB bandwidth)

PEAK DEVIATION: The lesser of 10 MHz or: n × Mod Rate where n = harmonic multiplication number (1 to 4). Refer to Frequency Ranges and Bandswitch Points Description.

**DEVIATION ACCURACY:** ± 10% (at 100 kHz rate)

SENSITIVITY: 1 MHz/Volt or 10 MHz/Volt

INPUT IMPEDANCE: Nominally 50 Ohms

# SWEEP TIME

# RANGE:

10 milliseconds to 200 seconds forward sweep times

Fastest possible sweep typically cycles once every 40 ms; fastest possible full band sweep typically cycles once every 150 ms.

MAXIMUM SPEED: Nominally 600 MHz/ms

**RESOLUTION:** Approximately 0.1% of current sweep time value.

ACCURACY: ±5% (sweeptimes ≤50 seconds)

Table 1-1. Model HP 8340B Specifications and Supplemental Performance Characteristics (13 of 15)

# INPUTS

PULSE MODULATION INPUT (Not applicable to HP 8340Bs equipped with Option 006, Delete Pulse Modulation) Front panel BNC female input connector. TTL compatible (open circuit is TTL high level and keeps RF on). Damage level is +12 Vdc, --20 Vdc.

# AMPLITUDE MODULATION INPUT

Front panel BNC female input connector. Nominal input impedance is 600 Ohms.

# FREQUENCY MODULATION INPUT

Front panel BNC female connector. Nominal input impedance is 50 Ohms. Full scale input voltage  $= \pm 1$  vac. Damage to internal circuitry will result if a signal with a peak voltage of  $\pm 8$  vac or greater is input.

# LEVELING EXT INPUT

Front panel BNC female input connector, Used for power meter leveling or crystal detector leveling. Input impedance in XTAL or METER modes is *nominally 1 MOhm*. Refer to EXTERNAL LEVELING specifications.

# FREQUENCY STANDARD EXT

Rear panel BNC female connector. Accepts 5 or 10 MHz signal from internal or external timebase. A BNC jumper connects this input to the HP 8340B's FREQUENCY STANDARD INT output for operation from HP 8340B's internal timebase. External signal input must be 5 MHz ±50 Hz or 10 MHz ±100 Hz, 0 to ±10 dBm. Nominal input impedance is 50 ohms.

# EXT TRIGGER INPUT

Rear panel BNC female connector. Triggers single sweep. Trigger signal must be >2 Vdc (10 Vdc max) and wider than 0.5 microseconds. *Nominal input impedence is 2 kOhms*.

# STOP SWP IN/OUT:

Rear panel BNC female connector. TTL high while sweeping, stops sweep when grounded externally. TTL low when HP 8340B stops sweep.

# HP 8755C INTERFACE

Rear panel. Connects via cable (HP Part No. 8120-3174) to HP 8755C Scalar Network Analyzer to provide Alternate Sweep function.

# HP 8410B INTERFACE

Rear panel 25-pin D-type connector. Permits multi-octave operation of HP 8410B/C Network Analyzer with HP 8340B via interface cable (HP Part No. 08410-60146). Also provides duplicates of these functions: Ext Trigger Input, Mute Output, Penlift Output, Neg Blank, and Z-Axis Blank/Mkrs. Also provides an input for a switch closure to execute the UP key function.

# **OUTPUTS**

# **RF OUTPUT**

Front panel Type N Female connector. Frequency output range is 10 MHz to 26.5 GHz. *Nominal output impedance is 50 Ohms.* SWR is shown in RF OUTPUT characteristics.

# SWEEP OUTPUT

Front and rear panel BNC female connectors. Supplies a voltage proportional to the sweep that ranges from approximately 0 Vdc (at start of sweep) to approximately  $\pm 10$  Vdc (at end of sweep), regardless of sweep width. In CW mode, the do voltage is proportional to percentage of full 10 MHz to 26.5 GHz range.

# 0.5V/GHz

Rear panel BNC female connector which outputs a voltage proportional to the instrument's output frequency (0.5V/GHz). Nominal load impedance should be greater than or equal to 4 KOhms. Accuracy of this signal is  $\pm 1\% \pm 2mV$ . This signal is intended for use with millimeter-wave source systems. This output can be changed to 1.0V/GHz (for use with the HP 6410C) by adding jumpers W1 and W2 on the A28 SYTM Driver board. The maximum output voltage of this signal is 19 vdc.

Table 1-1. Model HP 8340B Specifications and Supplemental Performance Characteristics (I4 of I5)

# OUTPUTS (Cont'd)

## **10 MHz REF OUTPUT**

Rear panel BNC (emale connector. Output power level is *nominally 0 dBm*, Output impedance is *nominally* 50 Ohms.

# FREQUENCY STANDARD INT

Rear panel BNC connector. Output frequency 10 MHz, output power nominally 3 dBm, 50 Ohm nominal output impedance. Jumpered to FREQUENCY STANDARD EXT for operation from HP 8340B's internal timebase.

#### MUTE OUTPUT

Rear panel BNC female connector. Mutes servo motor of X-Y recorder when the HP 8340B crosses a band switchpoint.

## PENLIFT OUTPUT J13

For operation with X-Y recorders. PENLIFT disables an X-Y recorder's ability to lower its pen during sweep retrace. If [SHIFT] [LINE] is pressed on the front panel, PENLIFT will also disable the pen during forward sweep band switchpoints. Because of X-Y recorder limitations PENLIFT will always disable the X-Y recorder's pen at sweep times under 5 seconds.

PENLIFT enables pen operation by providing a current path to ground for the X-Y recorder's pen solenoid. The voltage at the PENLIFT output in this state will be *approximately 0 Vdc*. Circuit impedance in this state is approximately 5 Ohms.

PENLIFT disables per operation by not providing a current path to ground for the X-Y recorder's pen solenoid. The voltage on the PENLIFT output will be equal to the X-Y recorder's pen solenoid supply voltage. Circuit impedance in this state is very high.

# NEG BLANK

Rear panel BNC female connector. Supplies negative rectangular pulse (approximately -5 Vdc into 2 kOhm load) during the retrace and band switchpoints of the RF output.

#### Z-AXIS BLANK/MKRS

Rear panel BNC female connector. Supplies positive rectangular pulse (approximately  $\pm 5$  Vdc into 2 kOhms) during the retrace and band switchpoints of the RF output. Also, supplies a  $\pm 5$  Vdc pulse when the RF is coincident with a marker frequency (intensity markers only).

#### AUX OUT

Rear panel Type-N female connector. Provides a 2.3 to 7.0 GHz fundamental oscillator output, nominally 0 dBm and 50 Ohm output impedance.

# REMOTE OPERATION

All functions (except line power) may be programmed via the Hewlett-Packard Interface Bus (HP-IB). Detailed Remote operation information is included in Section III, Operation.

# GENERAL

# ENVIRONMENTAL

Temperature: Operation at 0 to ~55°C, except as noted in electrical specifications.

Humidity: Passes 5 day cycling, +40°C, 95% relative humidity.

EMI: Controlled and radiated interference is within the requirements of CE03 and REO2 (relaxed by 10 dB) of MIL STD 461A, and within the requirements of VDE 0871/1978. Level 8 and CISPR publication 11 (1975).

Table 1-1. Model HP 8340B Specifications and Supplemental Performance Characteristics (15 of 15)

GENERAL (Cont'd)

# WARM-UP TIME

**Operation:** Requires 30-minute warmup from cold start, 0 to +55°C. Internal temperature equilibrium is reached after 2-hour warmup at stable outside temperature.

**Frequency Reference:** Reference time base is kept at operating temperature in STANDBY mode with the instrument connected to the ac power. For instruments disconnected from ac power for less than 24 hours, the aging rate is  $<1 \times 10^{-9}$ /day after a 72-hour warmup.

# POWER REQUIREMENTS

47.5 to 66 Hz; 100, 120, 220, or 240 volts (±10%); Typically, 500 VA maximum (40 VA in STANDEY).

# WEIGHT

Net Weight: 34 kg (75 lb)

Shipping Weight: 52 kg (112 lb)



Figure 6. Instrument Dimensions

# NOTES

- Maximum leveled power from 35°C to 55°C will typically be degraded from these specifications by no more than 2 dB.
- 2. Internally leveled, AM off. The POWER dBm display monitors that actual output power, giving accurate readings when unleveled, externally leveled, or when amplitude modulating with a signal that has a dc component. In these modes, the accuracy *typically degrades by =0.1 dB* over the tabulated values. The ENTRY DISPLAY shows the desired power level, or the desired external detector output voltage, exclusive of modulation.
- The ALC loop typically operates up to -20 dBm to enhance usability at those frequencies where leveled power greater than the maximum specified is available.