



# Agilent E5100A Network Analyzer

## Data Sheet

These specifications are the performance standards or limits against which the instrument is tested. When shipped from the factory, the E5100A meets the specifications listed in this section.

Values followed by (SPC) are supplemental performance characteristics.

### Source

#### Frequency characteristics

**Range** 10 kHz to 300 MHz

**Accuracy** (at  $23 \pm 5^\circ\text{C}$ ) .....  $\pm 20$  ppm  
With Option E5100A-1D5 (at 0 to  $55^\circ\text{C}$ , 20 minutes after power on) .....  $\pm 1$  ppm

**Stability** (at  $23 \pm 5^\circ\text{C}$ ) .....  $\pm 5 \times 10^{-6}/\text{day}$  (SPC)  
With Option E5100A-1D5 (48 hours after power on) .....  $\pm 2.5 \times 10^{-9}/8$  hours (SPC)

**Resolution** ..... 1 mHz

#### Output power characteristics

(measured at RF OUT 1, RF OUT 2 is terminated with  $50\ \Omega$  termination)

**Range** (nominal)

With Option E5100A-001 ..... -9 dBm to +11 dBm  
With Option E5100A-002 ..... -15 dBm to +5 dBm  
With Option E5100A-003 ..... -12 dBm to +8 dBm  
With Option E5100A-801 ..... -48 dBm to +22 dBm  
With Option E5100A-802 ..... -54 dBm to +16 dBm  
With Option E5100A-803 ..... -51 dBm to +19 dBm  
With Option E5100A-600 (at RF OUT 1) ..... -52 dBm to +18 dBm  
With Option E5100A-600 (at RF OUT 2) ..... -65 dBm to +5 dBm

**Resolution** ..... 0.1 dB

**Level accuracy** (at  $23 \pm 5^\circ\text{C}$ , 0 dBm output level, 50 MHz) .....  $\pm 1$  dB

**Flatness** (at  $23 \pm 5^\circ\text{C}$ , relative to 0 dBm output level at 50 MHz) ..... +2 dB, -4 dB  
With Option E5100A-803 ..... +2.5 dB, -4.5 dB



Agilent Technologies

With Option E5100A-801 or E5100A-802	
10 kHz ≤ frequency < 50 kHz .....	+1.5 dB, -6 dB (SPC)
50 kHz ≤ frequency ≤ 100 MHz .....	+2.5 dB, -4.5 dB
100 MHz < frequency ≤ 300 MHz .....	+3 dB, -5 dB
With Option E5100A-600	
10 kHz ≤ frequency < 50 kHz .....	+1.5 dB, -7 dB (SPC)
50 kHz ≤ frequency ≤ 100 MHz .....	+2.5 dB, -4.5 dB
100 MHz < frequency ≤ 300 MHz .....	+3 dB, -5 dB
<b>Linearity</b> (at 23 ±5 °C, relative to 0 dBm output level at 50 MHz) .....	±1 dB
With Option E5100A-801/802/803	
Maximum power level -70 dB ≤ power level	
< maximum power level -60 dB .....	±1.5 dB
Maximum power level -60 dB ≤ power level	
≤ maximum power level .....	±1 dB
<b>Power splitter</b>	
(When the analyzer is equipped with Option E5100A-001 or E5100A-003, delete this section.)	
<b>Insertion loss</b> (When the analyzer is equipped with Option E5100A-600, delete this item.) .....	6 dB (nominal)
<b>Output tracking</b>	
Without Option E5100A-600	
10 kHz ≤ frequency ≤ 100 MHz .....	0.1 dB (SPC)
100 MHz < frequency ≤ 300 MHz .....	0.2 dB (SPC)
With Option E5100A-600	
10 kHz ≤ frequency ≤ 100 MHz .....	13 dB ±0.3 dB (SPC)
100 MHz < frequency ≤ 300 MHz .....	13 dB ±0.5 dB (SPC)
<b>Equivalent output SWR</b>	
Without Option E5100A-600	
10 kHz ≤ frequency < 100 MHz .....	≤ 1.2 (SPC)
100 MHz ≤ frequency ≤ 300 MHz .....	≤ 1.4 (SPC)
With Option E5100A-600	
10 kHz ≤ frequency < 50 kHz .....	≤ 2.5 (SPC)
50 kHz ≤ frequency ≤ 100 MHz .....	≤ 1.2 (SPC)
100 MHz < frequency ≤ 300 MHz .....	≤ 1.4 (SPC)
<b>Spectral purity characteristics</b>	
<b>Non-harmonic spurious signals</b> (at < 300 MHz)	
With Option E5100A-001 (at -4 dBm output level) .....	< -45 dBc
With Option E5100A-002 (at -10 dBm output level) .....	< -45 dBc
With Option E5100A-003 (at -7 dBm output level) .....	< -45 dBc

With Option E5100A-600 (at 0 dBm output level) . . . . .	< -45 dBc
With Option E5100A-801 (at +6 dBm output level) . . . . .	< -45 dBc
With Option E5100A-802 (at 0 dBm output level) . . . . .	< -45 dBc
With Option E5100A-803 (at +3 dBm output level) . . . . .	< -45 dBc

**Phase noise** (at 10 kHz offset from 0 dBm fundamental) . . . . . < -90 dBc/Hz

#### Other source information

<b>Reverse power protection</b> . . . . .	20 dBm, 25 Vdc (SPC)
<b>Output connector</b> . . . . .	BNC female
<b>Output impedance</b> . . . . .	50 Ω (nominal)

#### Receiver

##### Input characteristics

<b>Frequency range</b> . . . . .	10 kHz to 300 MHz
1 MΩ input for Options E5100A-703/704/707/708 . . . . .	10 kHz to 5 MHz
<b>IF bandwidth (IF BW)</b> . . . . .	10 Hz to 30 kHz, 1, 1.5, 2, 3, 4, 5, 8 step (nominal)
<b>Impedance</b> . . . . .	50 Ω (nominal)
1 MΩ input for Option E5100A-703/704/707/708 . . . . .	1 MΩ // 30 pF (nominal)
<b>Return loss</b> (at 50 Ω input)	
10 kHz ≤ frequency < 100 MHz . . . . .	20 dB (SPC)
100 MHz ≤ frequency ≤ 300 MHz . . . . .	15 dB (SPC)

##### Maximum input level

50 Ω input

Frequency	RF attenuator	Maximum input level
10 kHz ≤ frequency < 200 kHz	25 dB	0 dBm
10 kHz ≤ frequency < 200 kHz	0 dB	-25 dBm
200 kHz ≤ frequency ≤ 300 MHz	25 dB	+5 dBm
200 kHz ≤ frequency ≤ 300 MHz	0 dB	-20 dBm

##### 1 MΩ Input for Options E5100A-705/706/707/708

Frequency <sup>1</sup>	RF attenuator	Maximum input level
10 kHz ≤ frequency < 200 kHz	25 dB	0.22 Vrms
10 kHz ≤ frequency < 200 kHz	0 dB	0.013 Vrms
200 kHz ≤ frequency ≤ 300 MHz	25 dB	0.40 Vrms
200 kHz ≤ frequency ≤ 300 MHz	0 dB	0.022 Vrms

1. Measurement frequency ≤ 5 MHz

**Damage level**

DC .....	25 Vdc
AC .....	20 dBm

**Averaging noise level** (at magnitude measurement, 23 ±5 °C, RF attenuator:0 dB, 50 Ω input)<sup>1</sup>

IF BW 30 kHz (at > 1 MHz) .....	-100 dBm
IF BW 10 kHz (at > 300 kHz) .....	-105 dBm
IF BW 3 kHz (at > 100 kHz) .....	-110 dBm
IF BW 1 kHz 30 kHz ≤ frequency < 100 kHz .....	-95 dBm
100 kHz ≤ frequency ≤ 300 MHz .....	-115 dBm
IF BW 300 Hz 10 kHz ≤ frequency < 100 kHz .....	-100 dBm
100 kHz ≤ frequency ≤ 300 MHz .....	-120 dBm
IF BW 100 Hz 10 kHz ≤ frequency < 100 kHz .....	-105 dBm
100 kHz ≤ frequency ≤ 300 MHz .....	-125 dBm

**Input crosstalk** (When the analyzer is equipped with Option E5100A-100,delete this section.) Reference input (0 dBm input level at 10 kHz to  
200 kHz and +5 dBm input level at 200 kHz to 300 MHz, RF attenuator:

25 dB, 50 Ω input)

Test input (RF attenuator: 0 dB, terminated with 50 Ω termination)

10 kHz ≤ frequency < 100 kHz .....	< -110 dB
100 kHz ≤ frequency ≤ 300 MHz .....	< -120 dB

**Source crosstalk**

(all RF OUT and input connectors are terminated with 50 Ω terminations)

Without Option E5100A-801/802/803 (at +5 dBm output level,  
RF attenuator: 0 dB, 50 Ω input)

10 kHz ≤ frequency < 100 kHz .....	< -110 dB (SPC)
100 kHz ≤ frequency < 250 MHz .....	< -125 dB (SPC)
250 MHz ≤ frequency ≤ 300 MHz .....	< -120 dB (SPC)

With Option E5100A-801/802/803 (at +16 dBm output level,

RF attenuator: 0 dB, 50 Ω input)

10 kHz ≤ frequency < 100 kHz .....	< -120 dB (SPC)
100 kHz ≤ frequency < 250 MHz .....	< -135 dB (SPC)
250 MHz ≤ frequency ≤ 300 MHz .....	< -130 dB (SPC)

---

1. When the analyzer frequency is identical to the transmitted interference signal frequency,  
refer to "EMC" in "general characteristics."

**Residual response**

(RF attenuator: 0 dB, except for the following points) ..... < -80 dBm  
50 kHz, 100 kHz, 95.825 MHz, 95.875 MHz, 159.791667 MHz, 159.825 MHz,  
159.841667 MHz, 159.875 MHz, 239.75 MHz, and 239.875 MHz

**Input connector** ..... BNC female  
With Option E5100A-705/706/707/708 ..... BNC female,  
Type-N female (for A, B inputs)

**Measurement mode**

With Option E5100A-100 ..... A  
With Option E5100A-200 or E5100A-600 ..... A/R, R/A, R, A  
With Option E5100A-300 ..... A/R, B/R, R/A, B/A, R/B, A/B, R, A, B  
With Option E5100A-400 ..... A/R, B/R, C/R, R/A, B/A, C/A, R/B, A/B,  
C/B, R/C, A/C, B/C, R, A, B, C

(When the measurement mode is either R/A, B/A, C/A, R/B, A/B, C/B,  
R/C, or A/C, the specification is SPC.)

**Magnitude characteristics****Absolute characteristics****Absolute amplitude accuracy**

(at  $23 \pm 5$  °C, -30 dBm input level for RF attenuator: 0 dB or -5 dBm input  
level for RF attenuator: 25 dB, 50 Ω input)  
±1 dB

**Ratio characteristics****Frequency response<sup>1</sup>**

(at  $23 \pm 5$  °C, -30 dBm input level for RF attenuator: 0 dB or -5 dBm input  
level for RF attenuator: 25 dB, the same RF attenuator setting for both inputs)  
50 Ω input

10 kHz ≤ frequency < 100 kHz ..... ±1 dB  
100 kHz ≤ frequency ≤ 100 MHz ..... ±0.5 dB  
100 MHz < frequency ≤ 300 MHz ..... ±1 dB

1 MΩ input for Option E5100A-703/704/707/708

(using 50 Ω feedthrough) ..... ±3 dB

---

1. Frequency response can be improved by calibration.

#### **Dynamic accuracy**

(at  $23 \pm 5^\circ\text{C}$ , 10 Hz IF BW,  $-10$  dBm reference input level relative to maximum input level,  $-20$  dBm test input level relative to maximum input level, except for ramp frequency sweep)

<b>Test channel input level RF attenuator</b>		<b>Dynamic accuracy frequency</b>	
25 dB	0 dB	Other	10 kHz to 50 kHz
+5 to $-5$ dBm <sup>1</sup>	$-20$ to $-30$ dBm <sup>2</sup>	$\pm 0.4$ dB	$\pm 0.4$ dB (SPC)
$-5$ to $-15$ dBm	$-30$ to $-40$ dBm	$\pm 0.09$ dB	$\pm 0.09$ dB (SPC)
$-15$ to $-45$ dBm	$-40$ to $-70$ dBm	$\pm 0.05$ dB	$\pm 0.05$ dB (SPC)
$-45$ to $-55$ dBm	$-70$ to $-80$ dBm	$\pm 0.06$ dB	$\pm 0.1$ dB (SPC)
$-55$ to $-65$ dBm	$-80$ to $-90$ dBm	$\pm 0.1$ dB	$\pm 0.3$ dB (SPC)
$-65$ to $-75$ dBm	$-90$ to $-100$ dBm	$\pm 0.3$ dB	$\pm 0.9$ dB (SPC)
$-75$ to $-85$ dBm	$-100$ to $-110$ dBm	$\pm 0.9$ dB	$\pm 3$ dB (SPC)
$-85$ to $-95$ dBm	$-110$ to $-120$ dBm	$\pm 3$ dB	N/A

With Option E5100A-100

(at  $23 \pm 5^\circ\text{C}$ , 10 Hz IF BW,  $-20$  dB input-A level relative to maximum input level, except for ramp frequency sweep, right after measuring reference)

<b>Test channel input level RF attenuator</b>		<b>Dynamic accuracy frequency</b>	
25 dB	0 dB	Other	10 kHz to 50 kHz
+5 to $-5$ dBm <sup>1</sup>	$-20$ to $-30$ dBm <sup>2</sup>	$\pm 0.4$ dB	$\pm 0.4$ dB (SPC)
$-5$ to $-45$ dBm	$-30$ to $-70$ dBm	$\pm 0.1$ dB	$\pm 0.1$ dB (SPC)
$-45$ to $-55$ dBm	$-70$ to $-80$ dBm	$\pm 0.1$ dB	$\pm 0.2$ dB (SPC)
$-55$ to $-65$ dBm	$-80$ to $-90$ dBm	$\pm 0.2$ dB	$\pm 0.6$ dB (SPC)
$-65$ to $-75$ dBm	$-90$ to $-100$ dBm	$\pm 0.6$ dB	$\pm 1.8$ dB (SPC)

#### **Trace noise**

(at 1 kHz IF BW, frequency  $> 305$  kHz,  $-5$  dBm input level for RF attenuator: 25 dB or  $-30$  dBm input level

for RF attenuator: 0 dB) .....  $< 10$  dBm rms

**Stability** .....  $0.02$  dB/ $^\circ\text{C}$  (SPC)

With Option E5100A-100

(at  $23 \pm 5^\circ\text{C}$ ) .....  $0.05$  dB/ $^\circ\text{C}$  (SPC)

---

1. 0 to  $-5$  dBm at 10 kHz to 200 kHz  
2.  $-25$  to  $-30$  dBm at 10 kHz to 200 kHz

### Phase characteristics

(When the analyzer is equipped with Option E5100A-100, delete this section.)

**Measurement mode** ..... Normal/Expanded

**Frequency response<sup>1</sup>**

(at 23 ±5 °C, -30 dBm input level for RF attenuator:

0 dB or -5 dBm input level for RF attenuator: 25 dB,

the same RF attenuator setting for both inputs, 50 Ω input)

10 kHz ≤ frequency < 100 kHz ..... ±5°

100 kHz ≤ frequency ≤ 100 MHz ..... +2.5°

100 MHz < frequency ≤ 300 MHz ..... ±5°

### Dynamic accuracy

(at 23 ±5 °C, 10 Hz IF BW, -10 dBm reference input level relative to maximum input level, -20 dBm test input level relative to maximum input level, except for ramp frequency sweep)

Test channel input level RF attenuator		Dynamic accuracy frequency	
25 dB	0 dB	Other	10 kHz to 50 kHz
+5 to -5 dBm <sup>1</sup>	-20 to -30 dBm <sup>2</sup>	±3°	±3° (SPC)
-5 to -15 dBm	-30 to -40 dBm	±0.6°	±0.6° (SPC)
-15 to -45 dBm	-40 to -70 dBm	±0.3°	±0.3° (SPC)
-45 to -55 dBm	-70 to -80 dBm	±0.3°	±0.6° (SPC)
-55 to -65 dBm	-80 to -90 dBm	±0.6°	±1.8° (SPC)
-65 to -75 dBm	-90 to -100 dBm	±1.8°	±6° (SPC)
-75 to -85 dBm	-100 to -110 dBm	±6°	±18° (SPC)
-85 to -95 dBm	-110 to -120 dBm	±18°	NIA

### Trace noise

(at 1 kHz IF BW, frequency > 305 kHz, -5 dBm input level for RF attenuator: 25 dB or -30 dBm input level for

RF attenuator: 0 dB) ..... < 50 mdeg rms

**Stability** ..... 0.15 deg/°C (SPC)

- 
1. This frequency response is only for the deviation from linear phase.  
Frequency response can be improved by calibration.
  2. 0 to -5 dBm at 10 kHz to 200 kHz
  3. -25 to -30 dBm at 10 kHz to 200 kHz

#### **Delay characteristics**

**Aperture frequency** ....  $\frac{200}{N-1}$ % to 100% of span, where  $N$  is number of points

#### **Accuracy** (at 23 ±5 °C, SPC)

In general, the following formula can be used to determine the accuracy, in seconds, of a specific group delay measurement:

$$\frac{\text{Phase accuracy} [\text{deg}]}{360 \text{ [deg]} \times \text{aperture} [\text{Hz}]} \text{ (sec)}$$

Depending on the aperture, input level, and device length, the phase accuracy used in either incremental phase accuracy or worst case phase accuracy.

### **General characteristics**

#### **Operating conditions**

When disk drive is in operation

Temperature ..... 10 to 40 °C

Humidity (at wet bulb ≤ 29 °C, without condensation) .. 15% ≤ RH ≤ 80%

When disk drive is not in operation

Temperature ..... 5 to 40 °C

Humidity (at wet bulb ≤ 29 °C, without condensation) .. 15% ≤ RH ≤ 80%

Altitude ..... 0 to 2,000 meters

Warm-up time ..... 30 minutes

#### **Non-operating conditions**

Temperature ..... - 20 to 60 °C

Humidity (at wet bulb ≤ 40 °C, without condensation) .. 15% ≤ RH ≤ 90%

Altitude ..... 0 to 4,572 meters

**Safety** ..... Certified by CSA-C22.2 No.1010.1-92, Based on IEC 1010-1 (1990)  
including Amendment 1 (1992)

**EMC<sup>1</sup>** ..... Complies with CISPR 11(1990)/EN 55011(1991): Group 1, Class A

Complies with IEC 801-2 (1991)/EN 55082-1(1992): 4 kV CD, 8 kV AD

Complies with IEC 801-3 (1984)/EN 55082-1(1992): 3 V/m

Complies with IEC 801-4 (1988)/EN 55082-1(1992): 1 kV power  
lines, 0.5 kV signal lines

**Power requirement** ..... 90 to 132 V or 198 to 264 V, 47 to 63 Hz, 400 VA max

**Weight** (depending on option) ..... 12 kg (SPC)

**Cabinet dimensions** ..... 425(W) x 177(H) x 425(D) mm (SPC)

---

1. When tested at 3 V/m according to IEC 801-3/1984, the averaging noise level will be within specifications over the full immunity test frequency range of 26 to 1000 MHz except when the analyzer frequency is identical to the transmitted interference signal test frequency.

## **Supplemental characteristics**

### **Measurement function**

<b>Number of measurement channels</b>	.....	1 to 4
<b>Display format</b>	.....	Cartesian
<b>Sweep parameter</b>	.....	frequency, power
<b>Sweep type</b>	E5100A .....	linear (step, ramp), list
<b>Measurement point per sweep</b>	E5100A .....	2 to 1,601

### **Others**

<b>Measurement calibration</b>	....	Response, response and isolation, 1-port 3-term
<b>Display</b>	.....	6.5 inch color LCD, 640 x 480 dots
<b>Flexible disk drive</b>	.....	720 Kbytes/1.2 Mbytes/1.44 Mbytes, DOS format, binary or ASCII format
<b>Flash disk</b>	.....	256 Kbytes
<b>Ram disk</b>	.....	256 Kbytes
<b>Programming</b>	.....	Instrument BASIC
<b>GPIB</b>	.....	ANSI/IEEE 488.2 compatible
<b>Parallel I/O port</b>	.....	16 bit output, 8 bit input/output, TTL level Option E5100A-005 ..... 8 bit output, 4 bit input, TTL level Option E5100A-006 ..... 16 bit output, 8 bit input/output, TTL level Option E5100A-007 ..... 16 bit output, 8 bit input, open collector, opto-isolated
<b>Printer</b>	.....	Parallel I/F (Centronics compatible), HP PCL
<b>Keyboard</b>	.....	mini-DIN (IBM PC compatible)
<b>External video monitor output</b>	.....	VGA

### **Connectors**

<b>Probe power</b>	... +15 V (300 mA max.), -12.6 V (160 mA max.), GND nominal (the maximum current values are total values of each probe connector)
--------------------	--

**EXT REF INPUT 10 MHz**

Frequency .....	10 MHz $\pm$ 5 ppm
Amplitude .....	0 $\pm$ 5 dBm (SPC)
Nominal impedance .....	50 $\Omega$

**REF OVEN (OPTION E5100A-1D5)**

Frequency (at 0 to 55 °C, 20 minutes after power ON) .....	10 MHz $\pm$ 1.0 ppm
Amplitude .....	2 $\pm$ 5 dBm (SPC)
Nominal impedance .....	50 $\Omega$

**INT REF OUTPUT**

Frequency (at 23 $\pm$ 5 °C) .....	10 MHz $\pm$ 20 ppm
Amplitude .....	0 $\pm$ 5 dBm (SPC)
Nominal impedance .....	50 $\Omega$

**EXT TRIGGER and EXT PROG RUN/CONT**

(Positive edge trigger)

$V_{ih}$ .....	+2 V to +5 V (SPC)
$V_{il}$ .....	0 V to +0.5 V (SPC)
Sink current ( $I_s$ ) .....	$I_s \leq 0.4$ mA (SPC)
Pulse width ( $T_p$ ) .....	$T_p \geq 20$ $\mu$ sec (SPC)

**Furnished accessories**

Accessories	Qty.	Agilent part number
Power cable	1	—
Sample program disk	1	E5100-180X0 <sup>1</sup>
CD-ROM (manuals)	1	E5100-905XX <sup>2</sup>
<b>Option E5100A-ABA add manuals</b>		
Function Reference	1	E5100-900X0 <sup>2</sup>
Programming Manual	1	E5100-900X7 <sup>2</sup>
User's Guide	1	E5100-900X1 <sup>2</sup>
Instrument BASIC Users Handbook	1	04155-90150
Instrument BASIC Users Handbook Supplement	1	E5100-900X5 <sup>2</sup>
<b>Option E5100A-0BW add Service Manual</b>		
Service Manual	1	E5100-901X0 <sup>2</sup>
<b>Option E5100A-1CM rack mount kit</b>		
Front handle kit	1	5062-3978
<b>Option E5100A-1CP front handle kit</b>		
Rack and handle kit	1	5062-3990
<b>Option E5100A-1CP rack and handle kit</b>		
Rack and handle kit	1	5062-3984
<b>Option E5100A-1D5 high stability frequency</b>		
BNC adapter	1	1250-1859
<b>Option E5100A-1F0 external keyboard</b>		
Keyboard	1	—

1. Furnished with special sample program disk (E5100-180X1) as well as the original one if Option E5100A-022/023 is designated. The number indicated by "X" in the part number of the sample program disk, is allocated for numbers increased by one each time a revision is made. The latest edition comes with the product.
- 2: The number indicated by "X" in the part number of each manual, is allocated for numbers increased by one each time a revision is made. The latest edition comes with the product.





## Agilent Email Updates

[www.agilent.com/find/emailupdates](http://www.agilent.com/find/emailupdates)

Get the latest information on the products and applications you select.



## Agilent Direct

[www.agilent.com/find/agilentdirect](http://www.agilent.com/find/agilentdirect)

Quickly choose and use your test equipment solutions with confidence.

## Agilent Open

[www.agilent.com/find/open](http://www.agilent.com/find/open)

Agilent Open simplifies the process of connecting and programming test systems to help engineers design, validate and manufacture electronic products. Agilent offers open connectivity for a broad range of system-ready instruments, open industry software, PC-standard I/O and global support, which are combined to more easily integrate test system development.

## Agilent Technologies' Test and Measurement Support, Services, and Assistance

Agilent Technologies aims to maximize the value you receive, while minimizing your risk and problems. We strive to ensure that you get the test and measurement capabilities you paid for and obtain the support you need. Our extensive support resources and services can help you choose the right Agilent products for your applications and apply them successfully. Every instrument and system we sell has a global warranty. Two concepts underlie Agilent's overall support policy: "Our Promise" and "Your Advantage."

### Our Promise

Our Promise means your Agilent test and measurement equipment will meet its advertised performance and functionality. When you are choosing new equipment, we will help you with product information, including realistic performance specifications and practical recommendations from experienced test engineers. When you receive your new Agilent equipment, we can help verify that it works properly and help with initial product operation.

### Your Advantage

Your Advantage means that Agilent offers a wide range of additional expert test and measurement services, which you can purchase according to your unique technical and business needs. Solve problems efficiently and gain a competitive edge by contracting with us for calibration, extra-cost upgrades, out-of-warranty repairs, and onsite education and training, as well as design, system integration, project management, and other professional engineering services. Experienced Agilent engineers and technicians worldwide can help you maximize your productivity, optimize the return on investment of your Agilent instruments and systems, and obtain dependable measurement accuracy for the life of those products.

[www.agilent.com](http://www.agilent.com)

For more information on Agilent Technologies' products, applications or services, please contact your local Agilent office. The complete list is available at:

[www.agilent.com/find/contactus](http://www.agilent.com/find/contactus)

### Phone or Fax

#### United States:

(tel) 800 829 4444  
(fax) 800 829 4433

#### Canada:

(tel) 877 894 4414  
(fax) 800 746 4866

#### China:

(tel) 800 810 0189  
(fax) 800 820 2816

#### Europe:

(tel) 31 20 547 2111

#### Japan:

(tel) (81) 426 56 7832  
(fax) (81) 426 56 7840

#### Korea:

(tel) (080) 769 0800  
(fax) (080) 769 0900

#### Latin America:

(tel) (305) 269 7500

#### Taiwan:

(tel) 0800 047 866  
(fax) 0800 286 331

#### Other Asia Pacific Countries:

(tel) (65) 6375 8100  
(fax) (65) 6755 0042  
Email: tm\_ap@agilent.com  
Contacts revised: 09/26/05

Product specifications and descriptions in this document subject to change without notice.

© Agilent Technologies, Inc. 2003, 2006  
Printed in USA, July 13, 2006  
5966-2888E



Agilent Technologies